

По вопросам продаж и поддержки обращайтесь:
Волгоград (844)278-03-48; Воронеж (473)204-51-73; Екатеринбург (343)384-55-89; Казань (843)206-01-48;
Краснодар (861)203-40-90; Красноярск (391)204-63-61; Москва (495)268-04-70; Нижний Новгород (831)429-08-12;
Новосибирск (383)227-86-73; Ростов-на-Дону (863)308-18-15; Самара (846)206-03-16; Санкт-Петербург (812)309-46-40;
Саратов (845)249-38-78; Уфа (347)229-48-12
Единый адрес: kmk@nt-rt.ru

www.kem.nt-rt.ru

Technical Datasheet



VTE* and VTE*-** (Ex)

Carrier-Frequency Pulse Amplifier

VTE* and VTE*-** (Ex) Carrier-frequency pulse amplifiers

Beschreibung

The integral carrier-frequency pickup of the VTE detects the r.p. of KEM flow meters. During this process the damping of a carrier-oscillator by the gears or blades is picked up through the meters body and evaluated. The frequency of the resultant amplitude modulation of the carrier is equal to the number of revolutions of the wheel and therefore a standard for the flow rate. The VTE may be operated in two- or three-wire mode and provides current or voltage squarewave pulses.

Types VTE * ** (Ex) are IS-approved according to ATEX 100a  II 2G EEx ia IIC T6. Our intrinsically safe barrier typ EWS is recommended to power the IS-approved versions.

Technical Data

Ex protection as per ATEX100a	 II 2 G EEx ia IIC T6)
Ambient temperature	–20 °C up to +50 °C
Medium temperature	max. +120 °C with a distance of at least 25 mm between flow meter and electronic housing max. +150 °C with a distance of at least 65 mm between flow meter and electronic housing
Supply voltage UB	7 up to 29 V DC, 8 up to 29 V DC for Ex versions
Quiescent current IR	< 4 mA
Frequency range	3 up to 3,000 Hz according to flow meter
Input impedance	< 100 Ω
Input	0.5 up to 500 mV
Electrical connection	3-pin terminals for supply and output signal, max. 2.5 mm ² cable gland 4–6 mm or 5-pin Amphenol plug type T3362500 for pin connections see page 4
Housing	aluminium, l = 64 mm, w = 58 mm, h = 38 mm
Pickup housing	stainless steel as per DIN 1.4104
Ingress protection	IP65 (DIN 40050)
Weight	approx. 250 up to 270 g
Outputs	frequency output, selectable: voltage level three-wire NPN/PNP a) three-wire active NPN high level: $U_{high} > U_B - 0.6 V - (2.6 k\Omega \cdot I_{out})$ low level: $U_{low} < 0.6 V + (1.3 k\Omega \cdot I_{out})$ b) three-wire passive NPN/open collector high level: $U_{high} > U - (1.3 k\Omega \cdot I_{out})$ low level: $U_{low} < 0.6 V + (1.3 k\Omega \cdot I_{out})$ U is the voltage applied at the output (open circuit voltage) max. 29 V c) three-wire active PNP (not for Ex versions) high level: $U_{high} > U_B - 0.6 V - (150 \Omega \cdot I_{out})$ low level: $U_{low} = \text{sperrnd}$ $I_{max.} = 60 \text{ mA}$; $P_{max.} \text{ an } R_s = 1 \text{ W}$; $R_s = 150 \Omega$ current level two-wire high level: $I_{high} > 2.2 \text{ mA}$ low level: $I_{low} < 1.4 \text{ mA}$

Safety-relevant parameters (only for Ex versions)

a) three-wire active NPN, version VTE*-3A

input:	KL1/KL2:	$U_{\max} = 30 \text{ V}$ $R_i = 1.2 \text{ k}\Omega$	$I_{\max} = 150 \text{ mA}$ $C_i = 0 \text{ nF}$	$L_i = 0$
output:	KL2/KL3:	$U_{\max} = 30 \text{ V}$ $R_i = 1.2 \text{ k}\Omega$	$I_{\max} = 25 \text{ mA}$ $C_i = 0 \text{ nF}$	$P_{\max} = 106 \text{ mW}$ $L_i = 0$


b) three-wire passive NPN/OC, version VTE*-3P

input:	KL1/KL2:	$U_{\max} = 30 \text{ V}$ $R_i = 1.2 \text{ k}\Omega$	$I_{\max} = 150 \text{ mA}$ $C_i = 0 \text{ nF}$	$L_i = 0$
output:	KL2/KL3:	$U_{\max} = 30 \text{ V}$ $R_i = 1.2 \text{ k}\Omega$	$I_{\max} = 500 \text{ mA}$ $C_i = 0 \text{ nF}$	$L_i = 0$

c) two-wire low power, version VTE*-2L

in-, output:	KL1/KL2:	$U_{\max} = 30 \text{ V}$ $C_i = 100 \text{ nF}$	$I_{\max} = 150 \text{ mA}$ $L_i = 0 \text{ nF}$	$P_{\max} = 175 \text{ mW}$ $R_i = 0$
	KL2/KL3	$U_{\max} = 30 \text{ V}$	$I_{\max} = 500 \text{ mA}$	
	KL3 n. c.			

Examples for connecting Ex versions




EWS = intrinsically safe power supply and separation amplifier

VTE* and VTE*-** (Ex) Carrier-frequency pulse amplifiers

Adjusting the output mode

The output mode is adjustable via jumpers located on the amplifier board. The table below is also printed on the inside of the housing top. With Ex-versions the output mode is adjusted by KEM according to customers' specifications and cannot be changed afterwards.

output mode	Jumper J3	Jumper J4	Jumper J5	Jumper J6
two-wire current level	off	on	off	off
three-wire active NPN	on	off	off	on
three-wire active PNP (PLC) ..	on	off	on	off
three-wire passive NPN	off	off	off	on




Electrical connection

The electrical connection is to be effected via one or two 3-pin terminals inside the amplifier which are accessible via cable sleeves 4–6 mm.




3-pin screw terminals

- 1 = +UB
- 2 = 0 V/GND
- 3 = Signal




5-pin plug

- 1 = +UB
- 2 = Signal
- 3 = 0 V/GND
- 4 = n. c.
- 5 = n. c.



Dimensional drawings (mm)



Ordering Information

VTE* - **

	K = short version with pickup for ZHM 02–04 and HM series
	L = long version with pickup for ZHM 02–07 and HM series
	R = short version with pickup for ZHM 01 and SRZ series
	S = long version with pickup for ZHM 01 and SRZ series up to +150°C (medium)
	Extension for electrical connection: »ST« 5-pin amphenol plug type T3362500 without extension: internal 3-pin screw terminals

VTE* - ** (Ex), Ex protection as per ATEX 100 II 2 G EEx ia IIC T6

	2L = two-wire current level low power
	3A = three-wire voltage level active NPN
	3P = three-wire voltage level passive NPN/OC
	K = short version with pickup for ZHM 02–04 and HM series
	L = long version with pickup for ZHM 02–07 and HM series
	R = short version with pickup for and SRZ series
	S = long version with pickup for ZHM 01 and SRZ series up to +150°C (medium)

Notes on Installation

The following has to be adhered to:

- Installation instructions for electrical devices
Installation instructions for associated intrinsically-safe devices
The »Special conditions for safe use« as per EC-Type Examination Certificate.
- The amplifier has to be installed in a way that the max. ambient temperature does under no circumstances exceed +50°C (consider self heating).
- With cables care should be taken, that the max inductivity and capacity of the respective voltage or gas group are not exceeded.
- Exceeding or falling below the regular measuring range will cause invalid frequency output signals.
- Shielded cables are to be used as connecting lines.
- Generally, supplied units have to be connected by an expert according to EMC stipulations.

Marking of the pulse amplifier

Two-wire connection

KEM Küppers Elektromechanik GmbH

 0123  II 2G Ex ia IIC T4

BVS 03 ATEX E 207

VTE*2* Nr. 12345678

-20 °C ≤ Ta ≤ 50 °C

KL 1/2 Ui = 30 V; li = 150 mA; Pi = 175 mW; Ci = 100 nF; Li = 0

KL 2/3 Ui = 30 V; li = 500 mA

KL 1 = Ub, 2 = 0 V; 3 = n.c.

Three-wire connection

KEM Küppers Elektromechanik GmbH

 0123  II 2G Ex ia IIC T4

BVS 03 ATEX E 207

VTE*-3* 12345678

-20 °C ≤ Ta ≤ 50 °C

KL 1/2 Ui = 30 V; li = 150 mA; Ri = 1.2 kΩ; Ci = 0; Li = 0

KL 2/3 Ui = 30 V; Ri = 1.2 kΩ; Ci = 0; Li = 0

VTE*-3A: li = 25 mA; Pi = 106 mW

VTE*-3P: li = 0.5 A

KL 1 = Ub; 2 = 0 V; 3 = output

The sticker indicates year of manufacture and person in charge of test.

По вопросам продаж и поддержки обращайтесь:

Волгоград (844)278-03-48; Воронеж (473)204-51-73; Екатеринбург (343)384-55-89; Казань (843)206-01-48;
Краснодар (861)203-40-90; Красноярск (391)204-63-61; Москва (495)268-04-70; Нижний Новгород (831)429-08-12;
Новосибирск (383)227-86-73; Ростов-на-Дону (863)308-18-15; Самара (846)206-03-16; Санкт-Петербург (812)309-46-40;
Саратов (845)249-38-78; Уфа (347)229-48-12
Единый адрес: kmk@nt-rt.ru